

Table 2: Measures of diagnostic accuracy				
	Total serum bilirubin		Total serum bilirubin	
	Rebound hyperbilirubinemia	NO Rebound hyperbilirubinemia	Rebound hyperbilirubinemia	NO Rebound hyperbilirubinemia
Test positive*	8	18	8	43
Test negative*	0	134	0	109
	Visual assessment		Transcutaneous bilirubinometry	
Sensitivity	100% (63.1-100)		100% (63.1-100)	
Specificity	88.2% (81.9-92.8)		71.7% (63.8-78.7)	
Positive predictive value	30.8% (14.3-51.8)		15.7% (7.0-28.6)	
Negative predictive value	100% (97.3-100)		100% (96.7-100)	
Positive likelihood ratio	8.44 (5.47-13.03)		3.53 (2.74-4.55)	
Negative likelihood ratio	0.00		0.00	
Diagnostic odds ratio	-		-	
ROC area	0.94 (0.92-0.97)		0.86 (0.82-0.89)	

*Refers to the test results of visual assessment (2nd and 3rd columns) or transcutaneous bilirubinometry (4th and 5th columns)